### National Institute of Plant Genome Research Arun Asaf Ali Marg, Post Box No. 10531, New Delhi – 110 067

### Technical - Bid

संख्या : ८/२०१८-१९/रा.पा.जी.अनु.सं/.एस एंड पी

दिनांक: 14/8/2018

विषय / Subject: मुहरबंद कोटेशन का निमंत्रण / Invitation of Sealed Quotations

Sealed Tenders are invited on behalf of Director, NIPGR from the Original Equipment Manufacturer or their authorized dealers for the purchase of **02 nos. of Autoclave** for the laboratory of our Institute, as per the following specifications in two bid system.

### **Technical Specifications**

- · Horizontal unit,
- Cylindrical with dimension: ~22"dia x 30"length,
- Capacity ~160 liters,
- Pressure Control Switch ranging from 10-20psi,
- Operated in dry and liquid cycles,
- Automatic shut-off at the end of both the sterilization and dry cycles,
- stainless steel chamber and door,
- Single hinged doors with high temperature resistant silicone gaskets,
- Double safety locking device to prevent door from opening while chamber is pressurized,
- Overheat cutoff to prevent burnout of heating elements in case of insufficient water in the chamber, working temperature thermostat to protect against overheating,
- overpressure release safety value,
- overheating and over temperature protector,
- Power Supply: 220-230VAC/50 Hz.

You are therefore requested to please send your offer in **two bid system** indicating the maximum discount offered, installation charges along with a copy of authorization certificate, issued by the Principals/OEM. The prices and taxes should be quoted separately in the Bid. The quotations must accompany a Demand Draft amounting to ₹ 6,800/- (Rupees Six Thousand Eight Hundred only), being the EMD in the name of Director, NIPGR, New Delhi and must be sent in a Sealed Envelope duly superscribed on top of envelope as "Quotation for 02 nos. of Autoclave" so as to reach to the undersigned latest by 05/9/2018 (3:00 PM), the same shall be opened on same day at 3.30 PM.

धन्यवाद.

(क्रय एवं भंडार अधिकारी)

Encl: Terms & Conditions (Annex - I)

# National Institute of Plant Genome Research Arun Asaf Ali Marg, Post Box No. 10531, New Delhi – 110 067

Annexure - I

## नियम और शर्तै:

- The tender document can be obtained from our Office on payment of ₹ 500/- (Rupees Five Hundred only) from 14/8/2018 to 04/9/2018 (01:00 P.M.). The tender document is also available on https://eprocure.gov.in and can also be downloaded free of
- Every tender shall be accompanied with the required Earnest Money Deposit in the form of Demand Draft drawn in favour of the "Director, NIPGR" payable at New Delhi. Any tender not accompanied by such earnest money will be rejected straight
- The rates quoted in the tender shall remain valid for a period 180 days from the date of issue of Award Letter. No tenderer can withdraw/or modify his tender or revoke the same within the said period. If a tenderer on his own withdraws or revokes the tender or revises or alters or modifies the tender for any item or condition within the period mentioned in the tender notice, his earnest money deposit shall stand forfeited. Notwithstanding foregoing, the Institute reserves the right to take other actions as deemed appropriate. In case the successful tenderer after award of acceptance of work fails to perform as per work order or violates any condition of tender, the security deposit/ performance security will be forefeited/encashed.
- NIPGR does not bind itself to accept the lowest or any tender and reserves the right to reject any or all tenders without
- In two-bid system the tenders should be submitted containing (a) Technical Bid consisting of all technical details along with commercial terms and conditions and (b) Financial Bid indicating item wise price for the items mentioned in technical bid. The technical bid and financial bid should be sealed by the bidder in separate covers duly super scribed and both these covers must be put in a bigger cover which should also be sealed and duly super scribed. NIPGR will not pay any expense, whatsoever incurred by tenderer for the preparation and submission of tenders.
- NIPGR will not pay any expense, whatsoever incurred by tenderer for the preparation and submission of tenders.
- The notice inviting tender, will form part of the contract agreement to be executed by the successful tenderer with the NIPGR.
- All the correspondence on the tender shall be addressed to the Director, NIPGR, Aruna Asaf Ali Marg, New Delhi and any communication addressed to anyone else shall not in any manner to be binding upon the NIPGR, Aruna Asaf Ali Marg. New
- The tenderer shall submit a copy of PAN/TIN/GST numbers allotted to them.
- NIPGR reserves the right to change the quantities of the units while issuing the letter of award of work.
- The successful tenderer shall be required to deposit an amount equal to 10% of the Tender value as Performance Security after adjusting the Earnest Money Deposit within 10 days from the date of issue of letter of acceptance. Performance Security may be deposited in the form of Demand Draft or Bank Guarantee from State Bank of India Or any Scheduled bank which shall be valid till completion of the warranty period of quoted / ordered items. Incase of violation of any condition of Tender, the Security Deposit / BG will be forfeited / revoked.
- The rates shall be inclusive of Transportation, loading, unloading, taxes etc., nothing extra will be paid.
- The supplier should be responsible for any damage and site clearance and nothing extra shall be paid.
- The bidder shall show the rate of GST and tax component in the financial bid separately. The GST shall be paid @ 5% as per GOI notification Nos. 45/2017 & 47/2017 dated 14/11/2017. The exemption certificate shall be provided at the time of placing Supply Orders.

(क्रय एवं भंडार अधिकारी)

# Arun Asaf Ali Marg, Post Box No. 10531, New Delhi - 110 067 National Institute of Plant Genome Research

PRICE BID

दिनांक: 14/8/2018

# संखाः ४/२०१४-१९/रा.पा.जी.अनु.सं/.एस एंड पी

Description	;	Foreign Currency (CIP)
Autoclave		
<ul> <li>Horizontal unit,</li> <li>Cylindrical with dimension: ~22"dia x 30"length,</li> </ul>		
<ul> <li>Capacity ~160 liters,</li> <li>Pressure Control Switch ranging from 10-20psi,</li> </ul>		
<ul> <li>Operated in dry and liquid cycles,</li> <li>Automatic shut-off at the end of both the sterilization and dry cycles,</li> </ul>		
<ul> <li>stainless steel chamber and door,</li> <li>Single hinged doors with high temperature resistant silicone gaskets,</li> <li>Single hinged doors with high temperature from opening while chamber is pressurized,</li> </ul>		
<ul> <li>Double safety locking device to prevent upon the promise.</li> <li>Overheat cutoff to prevent burnout of heating elements in case of insufficient water in the chamber,</li> </ul>		
working temperature thermostat to protect against overheating,		
<ul> <li>overpressure release salety value;</li> <li>overheating and over temperature protector;</li> </ul>		
• Power Supply: 220-230VAC/50 Hz.		